



PRODUCT RANGE

59-500-000 Durometer Shore 'A' Scale
59-500-111 Durometer Shore 'D' Scale

OPERATING PROCEDURE

Test Specimen Dimensions

SHORE **A**: 6mm thick minimum
SHORE **D**: 3mm thick minimum

Specimen size should allow measurement to be taken at least 12mm from any edge

The specimen surface should be flat and parallel to allow the presser face to contact the specimen over an area which has a minimum radius of 6mm from the durometers probe

The specimen may be constructed with layered pieces to achieve the necessary thickness requirement, however measurements taken on these specimens may not agree with those made on solid specimens, due to the surface faces between layers not being in complete contact.